

RECORD DATE : 12/08/2019

BATCH No: _____

CYCLE No: _____

UNIT		°C	bar		
TIME	EVENT	PV1	SV1	PV2	
16:25:28	LOG	LWL	121.0	0.00	
16:27:28	LOG	LWL	121.0	0.00	
16:29:28	LOG	50.0	121.0	0.00	
16:29:42	LOG	50.0	121.0	0.00	PROCESS START
16:31:42	LOG	111.0	121.0	0.48	
16:33:32	LOG	121.0	121.0	1.05	STERILISATION START
16:34:32	LOG	121.1	121.0	1.06	
16:35:32	LOG	121.1	121.0	1.06	
16:36:32	LOG	121.1	121.0	1.06	
16:37:32	LOG	121.1	121.0	1.06	
16:38:32	LOG	121.1	121.0	1.06	
16:39:32	LOG	121.1	121.0	1.06	
16:40:32	LOG	121.3	121.0	1.08	
16:41:32	LOG	121.3	121.0	1.08	
16:42:32	LOG	121.3	121.0	1.08	
16:43:32	LOG	121.3	121.0	1.08	STERILISATION END

FO (Mins) = 10.15

16:45:32	LOG	92.6	121.0	0.00	
16:47:32	LOG	59.4	121.0	0.00	
16:49:32	LOG	59.4	121.0	0.00	
16:51:32	LOG	59.4	121.0	0.00	
16:51:58	LOG	59.4	121.0	0.00	PROCESS START
16:53:58	LOG	78.4	121.0	0.00	
16:55:58	LOG	102.1	121.0	0.08	
16:57:45	LOG	121.0	121.0	1.04	STERILISATION START
16:58:45	LOG	121.1	121.0	1.06	
16:59:45	LOG	121.1	121.0	1.06	
17:00:45	LOG	121.1	121.0	1.06	
17:01:45	LOG	121.1	121.0	1.06	
17:02:45	LOG	121.1	121.0	1.06	
17:03:45	LOG	121.1	121.0	1.06	
17:04:45	LOG	121.1	121.0	1.06	
17:05:45	LOG	121.1	121.0	1.06	
17:06:45	LOG	121.1	121.0	1.06	
17:07:45	LOG	121.1	121.0	1.06	STERILISATION END

FO (Mins) = 9.97

17:09:45	LOG	101.4	121.0	0.01	
17:11:45	LOG	62.5	121.0	0.00	
17:13:45	LOG	62.5	121.0	0.00	
17:15:45	LOG	62.5	121.0	0.00	
17:17:45	LOG	62.5	121.0	0.00	